



**Questions & Answers referring to RFQ N° UNFPA/LKA/RFQ/19/07 based on questions received from suppliers by 3 pm on Friday, September 6, 2019 (Sri Lanka Time)**

**1. Can we quote for one item only?**

Yes, you may quote for any number of items in the list of medical equipment in the RFQ.

**2. What are the specifications for the spot lamps?**

Further specifications for the spot lamps are mentioned below:

Technical Specifications:

- Light mounted on mobile steel stand and base with 5 swivels
- Castors for medical examination
- Flexible arm, approx. 50-75 cm insulated, with on/off switch
- Voltage 220V/ 240V
- Bulb - Screw type 60 Watts or equivalent
- Adjustable reflector for positioning
- Free cord (approx. 3m) with 3 pin plug top, round pins
- Colour - Black, Brown or White
- Diameter of base approx. 25" (inches)
- Height (excluding flexible arm) 36" adjustable to 48"

Other conditions:

- Supplied with extra bulb (screw type)
- Clear instructions in English for use/ maintenance
- Diagrams for assembly and list of accessories/ parts
- One-year comprehensive warranty & warranty card

**3. We are expecting to quote for 50 Nos. of Mini Autoclaves, which has a capacity of 45L, though the requested capacity is 40L. Please let us know whether it is okay to quote the 45L Autoclave model.**

Yes, this is fine. If your company has the capacity to provide a competitive quotation, please go ahead and quote for this. In your quotation, please mention the specifications of supplies you are quoting for.

**4. What are the evaluation criteria for quotations?**

Quotations that meet the required specifications mentioned in the RFQ will be evaluated on their financial competitiveness.